

Insta-Test[®] ANALYTIC

DRINKING WATER 5



Free Chlorine, Total Chlorine, Total Hardness, pH & Nitrate
2 Test Strips


Code 2936-FP



LaMotte Company • Chestertown, Maryland 21620 USA • www.lamotte.com

INSTRUCTIONS:

Remove strip from foil. Immerse strip for 2 seconds. Remove with pads face up. SHAKE ONCE TO REMOVE EXCESS WATER.



Immediately read **Free Chlorine**, **Total Chlorine**, **Total Hardness** and **pH**. Wait 30 seconds, then read: **Nitrate**.

Free Chlorine ppm



0 0.5 1 3 5 10

Total Chlorine ppm



0 0.5 1 3 5 10

Total Hardness ppm (gpg)



0 50 100 200 400 ppm
(0) (3) (5.8) (11.7) (23) (gpg)

pH



4 5 6 7 8 9 10

Nitrate as NO₃-N ppm



0 5 10 25 50

Desiccant (Do Not Read)

